



**Product Code . EL-SM-11919**

## Expansion Index Test Apparatus

### Description

---

#### Expansion Index Test Apparatus

##### Specification:-

Construction: The unit consists of a cast aluminum base and collar, stainless steel specimen ring.

Base and Collar: Cast aluminum construction; brass hold-down rods.

Specimen Ring: Stainless steel; 101.6 x 25.4 mm (4" internal diam. x 1" h.).

Porous Stones: 101.4 x 12.7 mm (4" diam. x 1/2" thick).

Loading Weight: Brass; 5.7 kg (12.6 lbs.).




Dial Indicator: 0.3" range x 0.0001" graduations. 8 mm range x 0.002 mm graduations.

Test Duration: 24 hours or less on the cured sample.

```
{ "@context": "https://schema.org/", "@type": "Product", "name": "Expansion Index Test Apparatus",  
  "image": "http://www.educational-  
equipments.com/images/catalog/product/1675654126ExpansionIndexTestApparatusWithlogo.jpg",  
  "description": "Construction: The unit consists of a cast aluminum base and collar, stainless steel  
specimen ring. Base and Collar: Cast aluminum construction; brass hold-down rods. Specimen Ring:  
Stainless steel; 101.6 x 25.4 mm (4 internal diam. x 1 h.). Porous Stones: 101.4 x 12.7 mm (4 diam. x  
1/2 thick). Loading Weight: Brass; 5.7 kg (12.6 lbs.). Dial Indicator: 0.3 range x 0.0001 graduations. 8  
mm range x 0.002 mm graduations. Test Duration: 24 hours or less on the cured sample.", "brand":  
"Educational Lab Equipments", "sku": "5", "gtin8": "5", "gtin13": "5", "gtin14": "5", "mpn": "5",  
  "aggregateRating": { "@type": "AggregateRating", "ratingValue": "5", "bestRating": "5", "worstRating":
```

"0", "ratingCount": "15" } }

---

Educational Lab Equipments,  
#449, HSIIDC, Industrial Area, Saha, Haryana  
Direct Contact Details  +91-98173-19615  sales@educational-equipments.com  
 [www.educational-equipments.com](http://www.educational-equipments.com)